



Certificate of Compliance

Certificate: 2717186

Master Contract: 260038

Project: 2717186

Date Issued: June 10, 2014

Issued to: Walsn Meterage & Control

(Hebei) Ltd.
No.2 building,
Qinghua Technology Park
Langfang development Area,
Langfang, Hebei 065001
China
Attention: Will Zhang

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.



Leon Hu

Issued by: Leon Hu

PRODUCTS

CLASS 3631 06 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

CLASS 3631 86 - ELECTRICAL EQUIPMENT FOR MEASUREMENT USE - Certified to US Standards

Electromagnetic Flowmeter, Models SX series. Rated input 120 V ac, 20 W.

Notes:

Notes:

1. The above model is permanently connected, Equipment Class I, Pollution Degree 2, Measurement Category II.
2. Mode of operation: Continuous.
3. Environmental Conditions: Normal: -25 to 60 degree C, 2000 m max, humidity 80% rH maximum, non-condensing.



Certificate: 2717186

Master Contract: 260038

Project: 2717186

Date Issued: June 10, 2014

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 61010-1-04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

UL Std. No. 61010-1 (2nd Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements